

Towards Commercialization of Hydrophone Flow Shields to Improve Passive Acoustic Data Quality (CRADA 678) Final Report

March 2026

- 1 Emma Cotter (Pacific Northwest National Laboratory)
- 2 Ian Brownstein (XFlow Energy)
- 3 Ben Strom (XFlow Energy)



U.S. DEPARTMENT
of **ENERGY**

Prepared for the U.S. Department of Energy
under Contract DE-AC05-76RL01830

NOTICE

This report was produced by Battelle Memorial Institute under Contract No. DE-AC05-76RL01830 with the Department of Energy. During the period of commercialization or such other time period specified by the Department of Energy, the Government is granted for itself and others acting on its behalf a nonexclusive, paid-up, irrevocable worldwide license in this data to reproduce, prepare derivative works, and perform publicly and display publicly, by or on behalf of the Government. Subsequent to that period, the Government is granted for itself and others acting on its behalf a nonexclusive, paid-up, irrevocable worldwide license in this data to reproduce, prepare derivative works, distribute copies to the public, perform publicly and display publicly, and to permit others to do so. The specific term of the license can be identified by inquiry made to the Contractor or DOE. NEITHER THE UNITED STATES NOR THE UNITED STATES DEPARTMENT OF ENERGY, NOR ANY OF THEIR EMPLOYEES, MAKES ANY WARRANTY, EXPRESS OR IMPLIED, OR ASSUMES ANY LEGAL LIABILITY OR RESPONSIBILITY FOR THE ACCURACY, COMPLETENESS, OR USEFULNESS OF ANY DATA, APPARATUS, PRODUCT, OR PROCESS DISCLOSED, OR REPRESENTS THAT ITS USE WOULD NOT INFRINGE PRIVATELY OWNED RIGHTS.

Printed in the United States of America

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831-0062

www.osti.gov

ph: (865) 576-8401

fax: (865) 576-5728

email: reports@osti.gov

**Available to the public from the National Technical Information Service
5301 Shawnee Rd., Alexandria, VA 22312 ph: (800) 553-NTIS (6847)**

or (703) 605-6000

email: info@ntis.gov

Online ordering: <http://www.ntis.gov>

Towards Commercialization of Hydrophone Flow Shields to Improve Passive Acoustic Data Quality (CRADA 678) Final Report

March 2026

1 Emma Cotter (Pacific Northwest National Laboratory)
2 Ian Brownstein (XFlow Energy)
3 Ben Strom (XFlow Energy)

Prepared for
the U.S. Department of Energy
under Contract DE-AC05-76RL01830

Pacific Northwest National Laboratory
Richland, Washington 99354

CRADA Final Report Certification Memo

Memo Completion Date: 3/10/2026

CRADA Number: 678

CRADA Title: Towards Commercialization of Hydrophone Flow Shields to Improve Passive
Acoustic Data Quality

Responsible Technical Contact at DOE Lab (PNNL):

Emma Cotter

Provide a list of publications, conference papers, or other public releases of results, developed under this CRADA:

This project produced a report describing the results of flow shield testing for public distribution.

Provide a detailed list of all subject inventions, to include patent applications, copyrights, and trademarks:

No subject inventions were developed under this CRADA, but the work relates to a prior patent application: US Patent App. 18/900,269

Executive Summary of CRADA Work

This project performed testing, market research, and manufacturing refinement to ready patent-pending hydrophone flow shields developed at PNNL for commercialization. The flow shields reduce the impact of flow noise on underwater sound recordings in areas of strong flow, like tidal channels or rivers. Flow noise is a challenge in many applications where it is necessary to measure underwater sound, including measuring the sound produced by marine energy converters and monitoring marine mammals. This project addressed several key questions about flow shield performance that are high-priority for potential customers and refined existing manufacturing processes to be scalable for market-level production. Specifically, this project conducted calibrations and field evaluations of the flow shields to comprehensively evaluate their performance. Overall, the results indicate that the flow shields reduce flow noise by over 20 dB at frequencies below 50 Hz at flow speeds greater than 0.5 m/s and are proven to survive deployments more than 6 months long.

Summary of Research Results

Flow noise, or low frequency pseudo-sound caused by turbulent flow past a hydrophone, is a persistent challenge for passive acoustic measurements in locations with currents (tidal channels, rivers, ocean currents), in shallow water where wave orbital velocities can induce flow noise, and when hydrophones are mounted to mobile platforms (e.g., vessels). Flow shields can reduce the effects of flow noise, but there have been few quantitative evaluations of their performance. To address this gap, in 2024, Pacific Northwest National Laboratory (PNNL) conducted a side-by-side test of three flow shield designs and found that an oil filled, urethane enclosure manufactured by XFlow reduced flow noise by over 30 dB at frequencies below 40 Hz in a tidal channel with 1.3 m/s peak velocities and, unlike thin fabric flow shields, was likely sufficiently robust for long-term deployments (Cotter et al. 2024). While Cotter et al. (2024) demonstrated the potential of the oil-filled flow shield to improve acoustic measurements in dynamic environments, analysis was limited to frequencies below 20 kHz, and only one oil-filled flow shield was tested for a limited period of one week.

Therefore, PNNL and XFlow partnered to refine the design of the oil-filled flow shield and conduct further testing of its performance. This study addressed three key topics:

- High-frequency performance
- Additional field testing and evaluation
- Ease of assembly

Overall, the results indicate that the flow shields:

- Attenuate propagating sound by no more than 2 dB at frequencies below 30 kHz
- Reduce flow noise by over 20 dB below 50 Hz at flow speeds greater than 0.5 m/s
- Are proven to survive deployments more than 6 months long¹
- Are easy to assemble

¹ Note that failure was not observed at 6 months, rather this was the maximum test duration possible during the project period.

Pacific Northwest National Laboratory

902 Battelle Boulevard
P.O. Box 999
Richland, WA 99354

1-888-375-PNNL (7665)

www.pnnl.gov